

TERMOMETRU DIGITAL

fără contact cu infraroșu

PRODUS DE



OPERAREA PRODUSULUI



MODEL IR988




CARACTERISTICILE PRODUSULUI



MODEL IR988




Timp de răspuns
de 1 secundă




Cip inteligent



Senzor
inteligent



Memorare a 32
de valori măsurate




Alarmă la
temperatură
ridicăată



Infraroșu fără
contact



Ecran mare de
afișare




Alertă LED
retroiluminată
în 3 culori



Sunet de
alertă pentru febră



Mod °F / °C



CARACTERISTICI SPECIALE



Alertă LED retroiluminată în 3 culori

În modul pentru corpul uman, produsul poate afișa în mod inteligent rezultatele măsurătorilor și le poate afișa în diferite culori, după cum urmează.

32.0°C ~ 37.4°C (Verde)

37.5°C ~ 37.9°C (Portocaliu)


38.0°C ~ 42.9°C (Roșu)

Sunet de alertă pentru febră

Atunci când temperatura va depăși 38°C / 100,4°F, va emite un sunet pentru a informa că temperatura este anormală.



UTILIZAREA PRODUSULUI



Mulțimi



Supermarket



Spații publice



Companii & Instituții



Copii



Familie



Obiecte



Animale de companie

CERTIFICATELE PRODUSULUI





Attestation of Conformity

No. ICR Polska/  **CE**

Name and address of Registered Manufacturer: 

Product name: Non-contact infrared thermometer

Product type/model: IR988

Trade mark: N.A.

This Attestation confirms that the product meets the requirements of the following normative documents and within limits of its documents gives presumption of conformity with essential requirements of Directive 93/42/EEC.

Relevant EC Directive: Medical Device Directive 93/42/EEC

Conformity assessment procedure: EC Declaration of Conformity (Annex VII of Directive 93/42/EEC)

Classification: Class I according Rule 1 of Annex IX of Directive 93/42/EEC

Applied normative documents: EN 60601-1: 2006
EN 60601-1-2: 2015

Applied Quality Management System: n/a

This AoC will remain valid only if Quality Management System Certificate remains valid. The assessment process has been carried out in accordance with the program PC-P-07-07. Evaluation has been carried out in accordance with test report made by:

- CHINA CERREI (SICHUAN) LABORATORY

No. of test report: SC(20)-50165A-10-2-MDD

Issue date: 25.03.2020

Expiration date: 24.03.2025

The mutual obligations and rights of the certification are regulated by the contract No. ICR Polska/2020-3125.

This Attestation applies to products having the same attributes (parameters), intended use, that have been evaluated and meet the requirements of the aforementioned standard.



Director: Rafal Kalinowski

Warsaw, 25. 03. 2020.

ICR Polska Co. Ltd.
ul. Plac Przymierza 6, 03-944 Warszawa
www.icrpolska.com, e-mail: icrpolska@icrq.com







CERTIFICATE OF EMC COMPLIANCE

According to Directive: 2014/30/EU

Ref. No.: 

Applicant: 

Address: 

Manufacturer: 

Address: 

Product: Non-contact infrared thermometer

Brand Name: 

Model No.: IR988

The test sample of product has been passed, the test according to requirements of the following standards:

Standard(s): EN 60601-1-2:2007
EN 61000-3-2 :2006+A1:2009+A2:2009
EN 61000-3-3 :2013

Test report(s) No.: 


Based on the voluntary assessment of the product sample and technical file, we confirm that the above-mentioned product meets the requirements of the EC directive.
The CE mark as show below can be used, under the responsibility of the manufacturer or the importer, after completion of an EC declaration of conformity and compliance with all relevant EC directives.




Approved by/Date:  Mar. 11, 2020




Shenzhen ITC Product Testing Co., Ltd.
2nd Floor, Building A7, Xinhé Xinxing Third Industrial Area, Fuhai Road, Fuyong Street Bao'an District, Shenzhen, China
Tel: (86)-0755-33138690 Fax: (86)-0755-23071003 http://www.itclab.cn/





Shenzhen ITC Product Testing Co., Ltd. Report No. 

FCC TEST REPORT

Client Information:

Applicant: 

Applicant add: 

Brand Name: 

Product Information:

Product Name: Non-contact infrared thermometer


Model No.: IR988


Derivative model No.: N/A


Test Date: Mar. 06, 2020 to Mar. 13, 2020 Issue Date: Mar. 13, 2020


Test Standard: FCC Part 15, Subpart B: 2016
ANSI C63.4:2014

Test Result: PASS

Tested by: Amanda Chen 


Reviewed by: Apple Huang 

Approved by: John Liu 



This test report may be reproduced in full only
Test result presented in this test report is applicable to the tested sample only

CERTIFICATELE PRODUSULUI



Fiscal Year 2020

CERTIFICATION OF REGISTRATION

This certifies that:


Model No: IR988

Was registered with US Food and Drug Administration, Center for Devices and Radiological Health, pursuant to the Code of Federal Regulations 21 CFR 807, by Shenzhen Huacetong Testing and Certification Co., Ltd


Owner/Operator Number: [Redacted]
Device Listing#: See annex
Expiration Date: December. 31, 2020

Shenzhen Huacetong Testing and Certification Co., Ltd. will confirm that such registration remains effective upon request and presentation of this certificate until the end of the calendar year stated above, unless said registration is terminated after issuance of this certificate. Shenzhen Huacetong Testing and Certification Co., Ltd makes no other representations or warranties, nor does this certificate make any representations or warranties to any person or entity other than the named certificate holder, for whose sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate-holder's device or establishment by the U.S. Food and Drug Administration. Shenzhen Huacetong Testing and Certification Co., Ltd assumes no liability to any person or entity in connection with the foregoing.

Pursuant to 21 CFR 807.39, "Registration of a device establishment or assignment of a registration number does not in any way denote approval of the establishment or its products. Any representation that creates an impression of official approval because of registration or possession of a registration number is misleading and constitutes misbranding." The U.S. Food and Drug Administration does not issue a certificate of registration, nor does the U.S. Food and Drug Administration recognize a certificate of registration, Shenzhen Huacetong Testing and Certification Co., Ltd is not affiliated with the U.S. Food and Drug Administration.



Executive Director
Issued: Mar.30, 2020
Expiration Date: Dec.31, 2020



CERTIFICATE OF COMPLIANCE

2015/863/EU Restriction of Hazardous Substance
Ref. No.: [Redacted]

Applicant: [Redacted]
Address: [Redacted]

Manufacturer: [Redacted]
Address: [Redacted]

Product : Non-contact infrared thermometer
Brand Name: [Redacted]
Model No.: IR988

The test sample of product has been passed, the test according to requirements of the following standards:

Standard(s): IEC62321-3-1:2013;
IEC62321-4:2013+A1:2017;
IEC 62321-5:2013; IEC 62321-6:2015;
IEC 62321-7-1:2015; IEC 62321-7-2:2017;
IEC 62321-8:2017

Test report(s) No.: [Redacted]


The Certificate of compliance is based on a test procedure or an evaluation of the above-mentioned product. This is to certify that the above-mentioned product is in compliance with the EU RoHS Directive 2011/65/EU and its amendment directives 2015/863/EU (RoHS 2.0) of the European parliament.

ROHS Approved by/Date: [Signature] Mar. 13, 2020
Department Manager



Shenzhen ITC Product Testing Co., Ltd.
2nd Floor, Building A7, Xinhexin Third Industrial Area, Fuhai Road, Fuyong Street Bao'an District, Shenzhen, China
Tel: (86)-0755-33138690 Fax: (86)-0755-23071003 http://www.itclab.cn/

CERTIFICATE OF ACCREDITATION



Medical device quality management system

Registration No.: [Redacted]

We hereby certify that



The established medical device quality management system conforms to

ISO 13485:2016
Scope of certification/registration

Medical Devices, non-contact infrared temperature detector, electronic products, health measurement device research and development, production and sales

Issue date of this certificate: Mar 17, 2020 Valid date of this certificate: Mar 16, 2021
First registration: Mar 17, 2020 Registration deadline: Mar 16, 2023

(The effectiveness of this certificate shall be validated by periodic surveillance audit of ISP for maintenance. The intervals of not more than 12 months. Conditions on effectiveness of certificate see <http://www.isp-acpc.com>. Note: the use of outside mainland China)



Issued by: [Signature]

INSPEC INTERNATIONAL
www.isp-acpc.com

AMBALAJUL PRODUSULUI



Pentru oameni	Mulțimi nelimitate	Capacitate de memorare	32 de valori
Metodă de măsurare	Fără contact	Alimentare electrică	DC3V (AAA*2)
Zonă de măsurare	Frunte	Dimensiuni	150*81*41mm
Distanță de măsurare	3~5cm	Greutate	105.5g
Interval de măsurare	Corp: 34.8°C~42.5°C Obiect: 0°C~93°C (±1°C)	Eroare maximă admisibilă	35°C~42.0°C (±0.2°C) Others (±0.3°C)

Informații pachet	Cantitate: 50Bucăți/Cutie	Dimensiune cutie: 58*27*39.5cm
	Greutate brută: 8.2KG/Cutie	Volum cutie: 0.06186CBM

Tip container	20' GP	40' GP	40' HC/HQ
Număr cutii	533	1075	1233
Număr produse	26650	53750	61650
Volum (real)	32.97138	66.4995	76.27338

Bateriile nu sunt incluse în pachet