

DIGITAL, SEM CONTACTO

TERMÓMETRO INFRARROJO

POR

 LyfTrack™

 LyfTrack™



OPERACIÓN DE PRODUCTO



MODELO IR988




CARACTERÍSTICAS DEL PRODUCTO



MODELO IR988




En 1 Segundo



Chip Inteligente



Sensor Inteligente



Arreglo de memoria de 32 unidades




Alarma de alta temperatura



Infrarrojo sin contacto




Pantalla grande



Alerta LED retroiluminada de 3 colores



Tono Beep de Fiebre



Modo °F/°C



CARACTERÍSTICAS ESPECIALES



ALERTA LED RETROILUMINADA DE 3 COLORES

En el modo cuerpo humano, el producto puede mostrar de forma inteligente los resultados de la medición y solicitarlo con diferentes colores LED retroiluminados de la siguiente manera.

32.0°C ~ 37.4°C (Verde)

37.5°C ~ 37.9°C (Naranja)


38.0°C ~ 42.9°C (Rojo)

BEEP DE FIEBRE ALTA

Una vez que la temperatura supere los 38°C/100.4°F, emitirá el sonido para informar que la temperatura es anormal.



USO DEL PRODUCTO



Multitud



Supermercado



Lugares públicos



Empresas e Instituciones



Bebés



Familia




Objeto



Mascotas

CERTIFICADOS DE PRODUCTO



Fiscal Year 2020

CERTIFICATION OF REGISTRATION

This certifies that:


Model No: IR988

Was registered with US Food and Drug Administration, Center for Devices and Radiological Health, pursuant to the Code of Federal Regulations 21 CFR 807, by Shenzhen Huacetong Testing and Certification Co., Ltd


Owner/Operator Number: [Redacted]
Device Listing#: See annex
Expiration Date: December 31, 2020

Shenzhen Huacetong Testing and Certification Co., Ltd. will confirm that such registration remains effective upon request and presentation of this certificate until the end of the calendar year stated above, unless said registration is terminated after issuance of this certificate. Shenzhen Huacetong Testing and Certification Co., Ltd makes no other representations or warranties, nor does this certificate make any representations or warranties to any person or entity other than the named certificate holder, for whose sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate-holder's device or establishment by the U.S. Food and Drug Administration. Shenzhen Huacetong Testing and Certification Co., Ltd assumes no liability to any person or entity in connection with the foregoing.

Pursuant to 21 CFR 807.39, "Registration of a device establishment or assignment of a registration number does not in any way denote approval of the establishment or its products. Any representation that creates an impression of official approval because of registration or possession of a registration number is misleading and constitutes misbranding." The U.S. Food and Drug Administration does not issue a certificate of registration, nor does the U.S. Food and Drug Administration recognize a certificate of registration, Shenzhen Huacetong Testing and Certification Co., Ltd is not affiliated with the U.S. Food and Drug Administration.



Executive Director
Issued: Mar.30, 2020
Expiration Date: Dec.31, 2020



CERTIFICATE OF COMPLIANCE

2015/863/EU Restriction of Hazardous Substance
Ref. No.: [Redacted]

Applicant: [Redacted]
Address: [Redacted]

Manufacturer: [Redacted]
Address: [Redacted]

Product : Non-contact infrared thermometer
Brand Name: [Redacted]
Model No.: IR988

The test sample of product has been passed, the test according to requirements of the following standards:

Standard(s): IEC62321-3-1:2013;
IEC62321-4:2013+A1:2017;
IEC 62321-5:2013; IEC 62321-6:2015;
IEC 62321-7-1:2015; IEC 62321-7-2:2017;
IEC 62321-8:2017

Test report(s) No.: [Redacted]

The Certificate of compliance is based on a test procedure or an evaluation of the above-mentioned product. This is to certify that the above-mentioned product is in compliance with the EU RoHS Directive 2011/65/EU and its amendment directives 2015/863/EU (RoHS 2.0) of the European parliament.

ROHS Approved by/Date: [Signature] Mar. 13, 2020
Department Manager



Shenzhen ITC Product Testing Co., Ltd.
2nd Floor, Building A7, Xinxing Third Industrial Area, Fuhai Road, Fuyong Street Bao'an District, Shenzhen, China
Tel: (86)-0755-33138690 Fax: (86)-0755-23071003 http://www.itclab.cn/

CERTIFICATE OF ACCREDITATION



Medical device quality management system

Registration No.: [Redacted]

We hereby certify that

The established medical device quality management system conforms to

ISO 13485:2016
Scope of certification/registration

Medical Devices, non-contact infrared temperature detector, electronic products, health measurement device research and development, production and sales

Issue date of this certificate: Mar 17, 2020 Valid date of this certificate: Mar 16, 2021
First registration: Mar 17, 2020 Registration deadline: Mar 16, 2023

(The effectiveness of this certificate shall be validated by periodic surveillance audit of ISP for maintenance. The intervals of not more than 12 months. Conditions on effectiveness of certificate see <http://www.isp-acpc.com>. Note: the use of outside mainland China)



Issued by: [Signature]

INSPEC INTERNATIONAL
www.isp-acpc.com

EMPAQUE DE PRODUCTO



Para personas	Multitudes ilimitadas	Arreglo de memoria	32 unidades
Método de medición	Sin contacto	Suministro de energía	DC3V (AAA*2)
Sitio de medición	Frente	Tamaño	150*81*41mm
Distancia de medición	3~5cm	Peso	105.5g
Rango de medición	Cuerpo: 34.8°C~42.5°C Objeto: 0°C~93°C (±1°C)	Error máximo permitido	35°C~42.0°C (±0.2°C) Others (±0.3°C)

Información de paquete	Cantidad: 50Piezas/Cartón	Tamaño de cartón: 58*27*39.5cm
	Peso Bruto: 8.2KG/Cartón	Volumen de cartón: 0.06186CBM

Tipo de gabinete	20' GP	40' GP	40' HC/HQ
Cantidad de cartón	533	1075	1233
Cantidad de producto	26650	53750	61650
Volumen (Ajuste Actual)	32.97138	66.4995	76.27338

Baterías no incluidas en el paquete